

TENDER NOTICE

Sealed tenders subscribed as:

1. TENDER FOR purchase of Hospital furniture and machinery & equipments for the Super Specialty Block, Dr.RPGMC, HOSPITAL, KANGRA AT TANDA.

Tenders are invited by the Principal, Dr. R.P. Govt. Medical College, Tanda Distt. Kangra H.P. from the original manufacturers or their distributors so as to reach the Principal office Kangra at Tanda **on or before 45 days from the date of publication upto 1 PM** thereafter no tender will be entertained and the same will be opened on the same day at 2.00 PM in the presence of tenderers or their representatives. If that day happens to be holiday the tenders will be opened on next working day.

Firms should be registered with DGS & D Delhi or the Controller of Store H.P., Shimla. Tenderers should attach photocopies of the Registration letter with the tenders. The tender not quoted according to the laid down standard specification and conditions will be rejected straightway. The tender must be accompanied by earnest money of Rs.1,00,000/- (Rs. One Lac only) duly pledged in the name of the *PRINCIPAL, DR.R.P.GOV. MEDICAL COLLEGE, KANGRA AT TANDA (H.P.)* in the shape of fixed deposit receipt or bank draft. The tender without Earnest money or short of it or not in the form specified above will be rejected straightway. The earnest money of the approved tenderers will be returned after completion of supply as ordered. The Principal, Dr. R. P. Govt. Medical College, Kangra at Tanda reserves the right to reject any or all the offers without assigning any reason. The list of items to be purchased can be had along with terms and conditions of the supply from the office of Principal on any working day between 10.30 AM to 4.00 PM personally one day prior to final date of opening of tender on cash payment of Rs. 1000/- or Rs. 1100/- each by registered post (only M.O/TMO will be entertained) sufficiently in advance to avoid the risk of late dispatch of the tender forms. Delay in submission of tender form due to late receipt of MO/TMO will be the entire responsibility of the tenderers.

❖ Any dispute arising thereafter will be subject to the jurisdiction of District courts Dharamshala.

Note:-

- Tender specification and price bid should be quoted separately in sealed envelopes.
- Unsealed tenders/price bids will be rejected straight away.
- The tendered items should be F.O.R. Dr. R.P. Govt. Medical College, Kangra at Tanda.
- Conditional tenders will not be accepted at any cost.
- Tenders details can be have on the official web site www.rpgmc.ac.in

- sd-

Principal
Dr.R.P.Govt.Medical College
Kangra at Tanda (H.P).

PLEASE ENSURE-

1. ARTICLES ARE WITH WARRANTY OF ATLEAST FIVE YEARS. SUBSTANDARD ITEMS WILL NOT BE ACCEPTED.
2. THE METAL USED IS RUST FREE, EPOXY POWDER COATING WHERE EVER REQUIRED.
3. THE PAINT USED IS LEAD FREE
4. THE SPIRIT OF SPECIFICATIONS ARE TO BE FOLLOWED AS ULTIMATELY IT IS FOR THE NEEDY PATIENTS.
5. DEMONSTRATION AT DR.RPGMC TANDA IF REQUIRED BY THE DEPARTMENTS.

S. NO	ARTICLE	RELEVANT SPECIFICATIONS
1.	ICU BEDS ELECTRICALLY OPERABLE	RUST PROOF ELECTRICALLY OPERABLE TUCK AWAY OR SLIDING SIDERAILS ,FOUR SECTION MATTRESS BASE,WATER PROOF,12CM THICK MATRESS,SUITABLE TABLE AT FOOT END WHICH CAN COME OVER THE BED AND A CABINET AT HEAD END FOR KEEPING PATIENTS CHARTS AND MEDICINES,APPROPRIATE STOOL FOR PATIENTS ATTENDENT,DETACHABLE BED END,FACILITY FOR TRENDLENBERG AND ANTI TRENDLENBERG POSITION ISO CERTIFIED FDA/CE APPROVED WARRANTY FOR 5 YEAR
2.	ICU BEDS MANUAL OPERABLE	RUST PROOF MANUAL OPERABLE TUCK AWAY OR SLIDING SIDERAILS ,FOUR SECTION MATTRESS BASE,WATER PROOF,12CM THICK MATRESS,SUITABLE TABLE AT FOOT END WHICH CAN COME OVER THE BED AND A CABINET AT HEAD END FOR KEEPING PATIENTS CHARTS AND MEDICINES,APPROPRIATE STOOL FOR PATIENTS ATTENDENT,DETACHABLE BED END,FACILITY FOR TRENDLENBERG AND ANTI TRENDLENBERG POSITION ISO CERTIFIED FDA/CE APPROVED WARRANTY FOR 5 YEAR
3.	FOWLER BED WITH WHEEL WITH FOWLER MATTRESS	METALIC, HARD BED, ADJUSTABLE HEAD END, WITH ATTACHMENT FOR IV LINE STURDY WHEELS- DIAMETER 6 INCHES OR MORE WITH LOCKS, WITH NON PLASTIC WATERPROOF MATERIAL COVERED DURABLE MATTRESS,-12 CM THICK
4.	PATIENT SEMI-FOWLER BED WITH WHEEL WITH SEMI-FOWLER MATTRESS	METALIC, HARD BED, ADJUSTABLE HEAD END, WITH ATTACHMENT FOR IV LINE, STURDY WHEELS- DIAMETER 6 INCHES OR MORE WITH LOCKS, WITH NON PLASTIC WATERPROOF MATERIAL COVERED DURABLE MATTRESS, ,-12 CM THICK
5.	PATIENT BED WITH MATTRESS	METALIC, HARD BED, ADJUSTABLE HEAD END, WITH ATTACHMENT FOR IV LINE, STURDY WHEELS- DIAMETER 6 INCHES OR MORE WITH LOCKS, WITH NON PLASTIC WATERPROOF MATERIAL COVERED DURABLE MATTRESS, ,-12 CM THICK

6.	BED SIDE LOCKER	METALLIC, WITH DRAWER AND A LOCKABLE COMPARTMENT
7.	CARDIAC TABLE	METAL BODY, WOODEN TOP, STURDY TABLE, STANDARD SIZE
8.	BED SIDE STOOL	SOFT COVERING- WITH A TOP OF 2 FEET BY 2 FEET
9.	PATIENT LOCKER	METALIC, STURDY, 2 FEET BY 2 FEET BY 2 FEET, IN ALMIRAH CONFIGURATION, WITH FACILITY FOR LOCKING
10.	PATIENT WAITING CHAIRS	ERGONOMIC, METALLIC, STURDY , HEAVY, CAN BE FITTED IN THE GROUND,IN CLUSTER OF 3 – 5
11.	X RAY VIEW BOX	LED VIEW BOX WITH AUTOMATIC ON AND OFF
12.	DRESSING TROLLEY	DOUBLE DECKER, METALLIC, WITH WHEELS, STURDY
13.	WASTE DISPOSABLE SET(3X1)	SET OF THREE EACH, 100 LITER CAPACITY OR MORE, DURABLE PLASTIC,
14.	DUSTBINS SMALL(OFFICE WASTE)	METAL, CHROME, WITH DETACHABLE BIN FROM OUTSIDE SHELL
15.	OPD PATIENT STOOLS	METAL, STURDY, ADJUSTABLE TOP HEIGHT
16.	RACK FULL SIZE	METAL, 6 TIER, EACH TIER SEPARATED BY 2 FEET OR MORE, STURDY , HEAVY DUTY
17.	RACK HALF SIZE	METAL 3 TIER, EACH TIER SEPARATED BY 2 FEET OR MORE, STURDY , HEAVY DUTY
18.	EXAMINATION TABLE	METALLIC, WITH SOFT CUSHION TOP, STABLE, 6 FEET BY 2 FEET TOP, SHOULD NOT TOPPLE OVER IF PATIENT SITS ON ONE SIDE
19.	INJECTION TROLLEY	WITH MULTIPLE DRAWERS SUPPORTED BY BALL BEARINGS
20.	FOOT /STEP STOOL	2 STEPPED, METAL, WITH STABLE BASE
21.	WEIGHING MACHINE	FULLY CALIBRATED, HEAVY DUTY
22.	LINEN BED SHEETS	WHITE COLOUR, COTTON, WASHABLE,LENGTH 2.5 METERS ,BREADTH 1.5 METERS
23.	PILLOWS	WHITE COLOUR, COTTON, WASHABLE,LENGTH 0.5 METERS ,BREADTH 0.25 METERS

24.	PILLOW COVERS	WHITE COLOUR, COTTON, WASHABLE FOR THE ABOVE SR NO .39
25.	BLANKETS	WOOLEN, EASILY WASHABLE, STAIN FREE, FIT FOR HARSH WINTERS OF OUR STATE.
26.	PATIENT TROLLEY	LARGE WHEEL BASE, BRAKE LOCK, STABLE, DURABLE, HEAVY DUTY.
27.	IV STAND	WITH STABLE BASE
28.	BP APPRATUS ADULT	GOOD STANDARD, REPUTABLE MAKE, PROPERLY CALIBTATED
29.	BP APPRATUS NEONATE	GOOD STANDARD, REPUTABLE MAKE, PROPERLY CALIBTATED
30.	AMBU BAG ADULT	ADULT SIZE, SILICONE/ POLYCARBONATE BASED, TRANSPARENT WITH CARRY BAG, VOLUME 400ML OR MORE.
31.	AMBU BAG PEDIARIC	ADULT SIZE, SILICONE/ POLYCARBONATE BASED, TRANSPARENT WITH CARRY BAG, VOLUME 400ML OR MORE
32.	LARYNGOSCOPE	GOOD QUALITY, REPUTABLE USFDA APPROVED COMPANY AND PRODUCT, WITH 2 SPARE BULBS
33.	NEEDLE DESTROYER	ABS PLASTIC/ STAINLESS STEEL 304 GRADE INSULATED WITH CUTTING & TRIM PROPERTIES
34.	CHEETELS FORCEPS	STAINLESS STEEL OF 8 AND 10 INCHES SIZE, DULL AND NON GLARING
35.	CUT SECTION TRAY	WITH ALL PERTAINING ARTICLES,
36.	OXYGEN CYLINDER	LARGE IN CASE OF FALIURE OF CENTRAL SUPPLY
37.	WHEEL CHAIRS	STURDY, HEAVY DUTY, WITH BRAKES
38.	DRESSING DRUM	METALLIC, RUST PROOF

	LARGE	
39.	OT DRUM	STAINLESS STEEL OT DRUMS: ISO CERTIFIED, WARRANTY 5 YEARS, STANDARD SIZE FOR AUTOCLAVING OF OT GOWNS, SHEETS, LINEN, SEAMLESS & SIZES -12"X 15", 9"X 9" & 9"X 11"
40.	STEEL BOWL	LARGE BOWL FOR KEEPING SALINE IN OT PROCEDURES
41.	FOOT STEPS	2 STEPPED, METAL, WITH STABLE CASE
42.	PATIENTS SHIFTING HYDRAULIC TRALLY	OVERALL APPROX SIZE: 1800MM X 675MM X 750MM FRAME WORK MADE OF RECTANGULAR M.S. TUBES STAINLESS STEEL TOP IN 3 SECTIONS TRENDLENBURG POSITIONS ADJUSTABLE BY SCREW/PNEUMATIC GAS SPRING SYSTEM. BACKREST SECTION ADJUSTABLE ON RACHET LEG END SECTION CAN SLIDE UNDER THE MAIN SECTION HEIGHT ADJUSTABLE A PAIR OF KNEE CRUTCHES PRE TREATED AND EPOXY POWDER COATED.
43.	INSTRUMENT RACKS	RACKS (FULL SIZE) : METALLIC, 6 TIRES EACH SEPARATED BY 2 FEET OR MORE, STURDY, HEAVY DUTY
44.	GLASS DOOR ALMIRAH	GLASS DOOR ALMIRAH: METALLIC, FRONT GLASS, STYLE KEY LOCKS, STANDARD SIZE, MULTIPLE SHELVES, EPOXY COATING.
45.	SHOES RACK	STAINLESS STEEL, 4 SHELVES, STANDARD SIZE
46.	INSTRUMENT TROLLEY	-1080MML X 490MMW X 840MMH (APPROX) -25MM X 25MM SS SQUARE TUBE FRAME -TROLLEY MOUNTED AS 100MM DIAMETER CASTOR WHEELS -TWO SS SHELVESWITH RAILING ON TOP --SS BOWL 360MM DIAM, SS BUCKET 300MM DIAMETER
47.	SURGEONS CHAIR	REVOLVING BASE, MOUNTED ON CARTERS
48.	CRASH CART	CRASH CART : 1. UNIT MOVES ON THE FOUR CASTORS IN WHICH 2 ARE WITH BREAKS & 2 ARE WITHOUT BREAKS TOP & LOWER SHELVES OF THE UNIT MADE OF STAINLESS

		STEEL 6 NO. OF TOTAL DRAWERS PROVIDED ON THE UNIT, 3 ON EACH SIDE MADE UP OF PLASTIC & PLASTIC TRAY OF DIFFERENT COLORS ON THE FRAME TO KEEP EMERGENCY MEDICINES. TOP OF THE UNIT COVERED FROM ALL 3 SIDES WITH THE HELP OF STAINLESS STEEL ROD. THE HANDLE TO BE PROVIDED ON THE UNIT TO MOVE AROUND. THE LOWER SHELVE TO BE COVERED FROM ALL THE 3 SIDES WITH STAINLESS STEEL ROD FOR THE SAFETY OF MATERIAL SHOULD HAVE I/V RODS, O2 CYLINDER HOLDERS. ELECTRIC LEAD INCORPORATED IN THE UNIT.
49.	NEBULISER	ULTRASONIC , PIEZO ELECTRIC PRINCIPLE BASED, ISO CERTIFIED, FDA/CE APPROVED, WARRANTY 5 YEARS, COMPACT PORTABLE DESIGN, SUPPLIED WITH ALL ACCESSORIES & TUBING'S , AT LEAST 2 MASK TO BE SUPPLIED.
50.	STYLET	STYLETS (INTRODUCER, INTUBATING GUIDE) NON FRICTION COVERING. ENOUGH MALLEABILITY YET ENOUGH RIGIDITY TO MAINTAIN ITS SHAPE DURING INSERTION. RESISTANT TO CHIPPING AND BREAKING. DISTAL AND SMOOTH. MEANS TO LIMIT STYLETS ADVANCEMENT INTO THE TUBE. PROXIMAL END MAY HAVE ATTACHMENT THAT WILL FIT FIRMLY INTO THE TRACHEAL TUBE AND PREVENT ROTATION OF TUBE ON STYLET.
51.	FOGGER MACHINE/FUMIGATOR	FOGGING MACHINE : SHOULD BE SUITABLE FOR AIR FUMIGATION AND ENVIRONMENT DISINFECTION SHOULD BE SAFE FOR MEDICAL EQUIPMENTS SHOULD BE PORTABLE AND EASY TO CARRY AEROSOL SIZE IN SUBMICRON SHOULD HAVE STERILIZATION CAPACITY MINIMUM 1000 CUBIC FEET OR MORE SHOULD BE MADE UP OF HIGH STRENGTH NON CORROSIVE ENGINEERING PLASTIC AND STAINLESS STEEL ELECTRIC SHOCK PROOF AND UNBREAKABLE MOTOR HOUSING SHOULD HAVE TIMER FOR AUTO SWITCH OFF SHOULD BE SUPPLIED WITH 3 YEAR WARRANTY SHOULD WORK SATISFACTORY ON INDIAN POWER SUPPLY CONDITION (220 – 240V) THE COMPANY SHOULD PROVIDE 5 LITERS OF SUITABLE FOGGING SOLUTION WITH EACH UNIT
52.	FORMALIN CHAMBERS	FORMALIN CHAMBER : ISO CERTIFIED , WARRANTY 5 YEARS, STANDARD LARGE SIZE, SUITABLE FOR KEEPING THORACOSCOPIC/ BRONCHOSCOPIC INSTRUMENTS, MULTIPLE SHELVES, LOCK AT ONE OF THE END

53.	AIRWAYS	<p>GUEDEL OROPHARYNGEAL AIRWAYS:</p> <p>PLASTIC OR SILICON RUBBER MATERIAL. FIRM BUT A TRAUMATIC FINISH.</p> <ul style="list-style-type: none"> SIZES : 00, 0, 1, 2, 3 & 4.
54.	VAGINAL SPECULUM CUSCO SMALL MEDIUM LARGE	CUSCO'S SELF RETAINING BIVALVE VAGINAL SPECULUM
55.	SIM SPECULAM SMALL MEDIUM LARGE	STAINLESS STEEL RUST FREE
56.	DNC SET	COMPLETE STAINLESS STEEL RUST FREE GOOD GRIP ON INSTRUMENTS AND EASY USAGE CUSCO VAGINAL SPECULUM. SIMS VAGINAL SPECULUM MEDIUM. HEGAR UTERINE DILATORS SET OF 8 PIECES SIMS UTERINE SOUNDS MEDIUM BLAKE DOUBLE ENDED CURETTE WITH ONE SHARP AND ONE BLEND END 27 CM UTERINE FORCEPS 10" SPONGE HOLDING FORCEPS STRAIGHT 9 1/2 " JAWS 19 CM SPONGE HOLDING FORCEPS STRAIGHT 9 1/2 " SMALL JAWS VUSELLUM FORCEPS CURVED 9"(23 CM) OVUM FORCEPS 10" METAL CATHETERS 6" ,VAGINAL WALL RETRACTOR.
57.	NASAL SPECULUM	STAINLESS STEEL RUST FREE, STANDARD SIZE
58.	TONGUE DEPRESSOR	STAINLESS STEEL RUST FREE, STANDARD SIZE
59.	LARYNGEAL MIRROR MEDIUM &LARGE	STAINLESS STEEL RUST FREE, STANDARD SIZE
60.	NASOPHARYNEAL MIRROR	STAINLESS STEEL RUST FREE, STANDARD SIZE
61.	BULLS LAMP & HEAD MIRROR	STAINLESS STEEL RUST FREE, STANDARD SIZE
62.	STERILIZER	STAINLESS STEEL RUST FREE, STANDARD SIZE
63.	STRAIGHT NEEDLE HOLDING FORCEPS	STAINLESS STEEL RUST FREE, STANDARD SIZE
64.	KOCHERS FROCEPS WITH TOOTHED JAW	STAINLESS STEEL RUST FREE, STANDARD SIZE

65.	ALLIS TISSUE FORCEPS	STAINLESS STEEL RUST FREE, STANDARD SIZE
66.	LONG STRAIGHT HEMOSTATIC FORCEPS	STAINLESS STEEL RUST FREE, STANDARD SIZE
67.	KIDNEY TRAY MEDIUM & LARGE	STAINLESS STEEL RUST FREE, STANDARD SIZE
68.	SUCTION MACHINE	NOISELESS OPERATION AND HAVE THE EFFICIENCY TO BUILD INSTANT VACUUM DUE TO OIL IMMERSSED PISTON PUMP. IT IS USED TO SUCTION OUT THE SALIVA OR PHLEGM AND REMOVES EXCESSIVE SECRETIONS AND KEEPS THE MOUTH AND AIRWAY CLEAR. WE OFFER SPECIALLY DESIGNED SELF-SEALING RUBBER CAPS IN THESE MACHINES AND ARE INCORPORATED WITH MOTOR CONNECTED TO A HIGHLY EFFICIENT ALL IMMERSSED PUMP MECHANISM.
69.	PATIENTS BED SIDE MONITORS	WITH PROVISION FOR HR ,ECG,O2 SATURATION,NIVP, RR, TEMP.AND WALLMOUNTED ,LCD,TOUCH SCREEN
70.	BED PAN MALE/FEMALES	LEAK PROOF ,STAINLESS STEEL(304 GAUGE 18)WITH ARGON WELDING CAPACITY IL
71.	URINE POTS	LEAK PROOF ,STAINLESS STEEL(304 GAUGE 18)WITH ARGON WELDING CAPACITY IL
72.	SUTURE BOX	ISO CERTIFIED ,WARRENTY 5 YRS STANDERED SIZE ,FOR KEEPING SUTURES AND DISPOABLE FOR EMERGENCY CASES COMPCT PORTABLE DESIGN,MULTIPLE SH ELVES/DIVISION FOR DIFFERENT SUTURES
73.	CIDEX CONTAINER	ISO CERTIFIED ,WARRENTY 5 YRS STANDERED LARGE SIZE ,C ORROSIVE RESISTANT
74.	GLUCOMETER	MEASURING RANGE: 10-600 MG/DL , SAMPLE SIZE: 1-2 μ L (1 μ L (MICROLITRE) - 1 THOUSANDTH OF A MILLILITER) MEASURING TIME: APPROX 5 SEC IF TEST STRIP IS IN METER WHEN APPLIED APPROX 10 SEC IF TEST STRIP IS OUTSIDE THE METER WHEN APPLIED POWER SUPPLY WITH BATTERY WITH DURABLE BATTERY LIFE: APPROX. 1000 TESTS OR APPROX 1 YEAR. AUTOMATIC POWER-OFF: AFTER 30 OR 90 SECONDS DEPENDING ON OPERATING

		STATUS. MEMORY: 350 TEST RESULTS WITH TIME AND DATE, AS WELL AS 7,14 AND 30 DAY AVERAGE
75.	STETH OSCOPE ADULT	STANDARD ADULT STEPHOSCOPE HAVING BOTH DIAPHROGM AND BELL FUNCITON.
76.	STETH OSCOPE PEDIATRIC	STNADARD PEDIATRIC STETHOSCOPE HAVING BOTH DIAPHRAGM AND BELL FUNCITION.
77.	SCALE BABY ELECTRONIC	POLYCARBONATE PAN, CAND WEIGH UPTO 10 KG, LARGE DISPLAY.
78.	MAYO TRALLY OVER OT TABLES	FOR C.T.V.S O.T, S.S TRAY/ TOP SIZE 910 MM×1200 MM×500 MM APPROX MOUNTED ON FOUR CASPRI OR MORE WITH BREAKS.
79.	LARGE INSTRUMENT TRALLY OVER OT TABLES	FOR C.T.V.S O.T, S.S TRAG/ TOP SIZE APPROX. 1400 MM×1200 MM SHOULD BE MOUNTED ON 50 MM CASPRI OR MORE WITH BREAKS.
80.	CURVED INSTRUMENT TRALLY	FOR C.T.V.S O.T, S.S TRAG/ TOP SIZE 1000×400×800 MM APPROX, MOUNTED ON CASPRI WITH BREAKS.
81.	DVT PUMPS	CONTROL UNIT TO REGULATE G, GRANTED, PRESSIVE, TO BOTHE LOWER LIMBS CAN BE ATTCHED TO BED, O.T TABLE.
82.	PULSE OXYMETRE	PORTABLE ,TABLE MOUNTED, ELECTRICTY OPERTATED, WITH PROVISION FOR MEASURING SPO2 AND HEART RATE DURING ENDOSCOPY/ERCp